

| TECH | NICAL | SHEET |
|------|-------|-------|
| | | |

COMPANY WITH QUALITY SYSTEM CERTIFIED BY DNV GL SYSTEM CEA Sir Safety System S.p.A. Single-member company VAT Nr. IT 03359340548 export@sirsafety.com Wholly-owned subsidiary of Sir Holding Srl Viale Europa 74 www.sirsafety.com = ISO 9001 = Via dei Fornaciai, 9 - Santa Maria degli Angeli 20090 Cusago Milano DNV.GL 06081 Assisi - Perugia - Italy Tel. +39 02.90394575 (R.A.) Tel.+39 075.8043737 (R.A.) - Fax +39 075.8048686 Fax +39 02.90394441 ISO 9001 Name: Gas Alert BW CLIP H2S - 2 Years Code: 81516N Gas Detectors - BW Description The gas detector BW Clip H2S is ideal for those who perform activities in hazardous environments. It operates up to three years maintenance-free. Just turn on the device and it runs continuously - no need for calibration, sensor replacement, battery replacement or battery charging. With BW Clip for H2S- 2 year it is possible to put the device in hibernation mode, by placing it in an appropriate housing when not in use for more than a week, thus it extends its life by that period of time. CO It features: - Maintenance-free: no sensor or battery changes necessary. - Compact, lightweight design with one-button operation. - Designed to withstand harsh environments and extreme temperatures. - Automated self-test of battery, sensor and electronics. Wide-angle flash, which alerts simultaneously with audible and vibrating alarm. - Automatic logging of the 35 most recent gas events and bump test results. Compact 4,1 x 5,0 x 8,7 cm Lightweight: 92 g Humidity:5% - 95% UR (non-condensing) Visual, vibrating and acoustic alarm (95 dB); low/high. Test: Activated detectors automatically perform one internal diagnostic test every 24 hours. Ingress Protection: IP 66/67

| Category | Colours | Available Sizes |
|-----------------|--|--|
| III | Yellow | |
| Quantity/Carton | Important Information | Recommended for |
| 1 | Make sure the product is in good condition, undamaged and clean before use. Once you become aware of a damage, you should immediately replace it. | Suitable to detect the presence of Sulphuric Acid. |
| Dexterity | EC Standards | Storage and cleaning Keep in a cool dry place, far from heat sources. |