



Sir Safety System S.p.A. Single-member company Wholly-owned subsidiary of Sir Holding Srl Via dei Fornaciai, 9 - Santa Maria degli Angeli 06081 Assisi - Perugia - Italy Tel.+39 075.8043737 (R.A.) - Fax +39 075.8048686 VAT Nr. IT 03359340548 Viale Europa 74 20090 Cusago Milano Tel. +39 02.90394575 (R.A.) Fax +39 02.90394441 export@sirsafety.com www.sirsafety.com

COMPANY WITH QUALITY SYSTEM CERTIFIED BY DNV GL = ISO 9001 =



Name: Gas Alert Extreme CLO2 Code: 81577 Gas Detectors - BW

Description

Gas Alert Extreme CLO2 features:

- One sensor.
- Full function self-test of sensor, battery status, circuit integrity and audio/visual alarms. It constantly shows the gas concentration on the display.
- Low, high, TWA and STEL alarms with adjustable settings.
- Acoustic (95dB), visual and vibrating alarm. Impact and water resistant casing. Also available in data logging version.
- Compact 2.8 x 5.0 x 9.5 cm.
- Lightweight 82 g.

Gas detector: Chlorine Dioxide.

The product is designed and manufactured to comply with the essential health and safety requirements of Directive 89/686/EEC, and to conform to the requirements of EN 60079.

It is classified as PPE Cat. III.



| Category III | Colours Yellow | Available Sizes |
|--------------|---|--|
| | | |
| 1 | Make sure the product is in good condition, undamaged and clean before use. Once you become aware of a damage, you should immediately replace it. | Suitable to detect the presence of Chlorine Dioxide. |
| Dexterity | EC Standards | Storage and cleaning |
| | EN 60079-0 | Keep in a cool dry place, far from heat sources. |
| | EN 60079-11 | |
| | EN 60079-26 | |
| | | |