

Name: Gas Alert Microclip X3 2 Gas (H2S, CO)

Code: 81607X3

Gas Detectors - BW

Description

Gas Alert Microclip X3 2 Gas, it detects 2 gases (Sulphuric Acid, Carbon Monoxide).

It features a protection level equivalent to IP68, with water- resistance up to 45 minutes at a depth of 1.2 m. Battery operating time up to 18 hour, single button operating. Reliable battery performance, Reflex[™] technology to ensure the reactivity of the sensor, compact, lightweight and easy to use, green IntelliFlash[™] flashing light for an immediate visual confirmation.

Five-year life expectancy — plus a three-year warranty.

An all-new oxygen sensor that operates consistently for five years — with no replacement costs for the life of the detector.

It features:

- Functioning guaranteed for at least a 12 hour shift, down to -20 ° C,
- Rechargeable lithium polymer long-lasting battery, tested to withstand extreme conditions,
- Ergonomic, lightweight and easy to wear,
- Gas concentration constantly displayed on LCD screen,
- Alarm lighbars with a wide viewing angle,
- The IP68 protection ensures maximum protection against dust and water: up to 45 minutes at a depth of 1.2 m,
- Simple operation, tamper-proof with a single button and intuitive LCD icons,
- IntelliFlash[™] light shows whether the detector has been subjected to impact tests and is working properly,
- Full function self-test of sensor, battery status, circuit integrity and audio/visual alarms.

Dimensions: 11,3 x 6,0 x 3,2 cm;

Weight: 179 g

Temperature: from - 20°C up to 50 °C.

Humidity: from 0 up to 95% RH (without moisture)

Acoustic (95dB), visual and vibrating alarm. Low, high, TWA, STEL and OL (Over Limit).

Battery operating time: 18 hours; It recharges in less than 6 hours

User options:

Reliability beep; Set STEL interval; On/off sensor; Alarms with manual reset; Safe display; stealth mode; calibration block; Auto-zero at startup; Automatic O2 calibration at startup; Set bump test interval; Mandatory impact test to overcoming of expiring date.

IP68 Nominal values: up to 45 minutes at a depth of 1,2 m.

Certifications: Class I, Div. 1, Group A, B, C, D

ATEX: II 1 G

Ex ia IIC T4 Ga

IECEX: Ex ia IIC T4 Ga

European Compliance

KTL



Category

III

Colours

Yellow

Available Sizes

Quantity/Carton

Important Information

Recommended for

Sir Safety System S.p.A. Single-member company
Wholly-owned subsidiary of Sir Holding Srl
Via dei Fornaciai, 9 - Santa Maria degli Angeli
06081 Assisi - Perugia - Italy
Tel. +39 075.8043737 (R.A.) - Fax +39 075.8048686

VAT Nr. IT 03359340548
Viale Europa 74
20090 Cusago Milano
Tel. +39 02.90394575 (R.A.)
Fax +39 02.90394441

export@sirsafety.com
www.sirsafety.com

COMPANY WITH
QUALITY SYSTEM
CERTIFIED BY DNV GL
= ISO 9001 =



TECHNICAL SHEET

| | | |
|-----------|---|---|
| 1 | Make sure the product is in good condition, undamaged and clean before use. Once you become aware of a damage, you should immediately replace it. | Suitable to detect the presence of Sulphuric Acid and Carbon Monoxide. |
| Dexterity | <p>EC Standards</p> <p>EN 60079-0</p> <p>EN 60079-11</p> | <p>Storage and cleaning</p> <p>Keep in a cool dry place, far from heat sources.</p> |



Sir Safety System S.p.A. Single-member company
Wholly-owned subsidiary of Sir Holding Srl
Via dei Fornaciai, 9 - Santa Maria degli Angeli
06081 Assisi - Perugia - Italy
Tel. +39 075.8043737 (R.A.) - Fax +39 075.8048686

VAT Nr. IT 03359340548
Viale Europa 74
20090 Cusago Milano
Tel. +39 02.90394575 (R.A.)
Fax +39 02.90394441

TECHNICAL SHEET

export@sirsafety.com
www.sirsafety.com

COMPANY WITH
QUALITY SYSTEM
CERTIFIED BY DNV GL
= ISO 9001 =

